



Ulrich Métrologie inc.
Ulrich Metrology Inc.
 9912, Côte-de-Liesse
 Montréal (Québec) H8T 1A1


Tél. (514) 631-6653
 Fax (514) 631-6122
 info@ulrich.ca
 www.ulrich.ca

CALIBRATION CERTIFICATE

Certificate no.: 330563
Identification: WMT001
Description: SCALE, GLASS MOTIC
Manufacturer: MOTIC

Calibration date: September 19, 2012
Certificate issued: September 19, 2012
Interval: 12 months
Due date: September 19, 2013
Procedure no.: CP-03 rev. 2
Procedure date: 2010-07-08
Environment: CLAS Type 1 Laboratory
Temperature: 20 ± 1°C
Humidity: 35 - 55% RH
Metrologist: XCH

Property of: G&R COLD FORGING PLANT 3
 7085 SMITH INDUSTRIAL DRIVE
 MCGREGOR, ON N0R 1J0

Approved by: 
 David Llorens, Quality Manager

This calibration certificate is issued in accordance with the applicable requirements of ISO/IEC 17025 and Ulrich Metrology's quality manual QM-08. Measurement results provided are traceable to either the National Research Council Canada (NRC), the National Institute of Standards and Technology (NIST), a national laboratory of another country signatory to the CIPM Mutual Recognition Arrangement (MRA), or a calibration laboratory accredited by an accrediting body with which Canada has an equivalence agreement.

CALIBRATION STANDARDS

Identification	Description	Cal. date	Due date
301	Measuring Machine, Sip 550m	2012-02-22	2013-02-28

MEASUREMENT UNCERTAINTY

± 0.0013 mm.

The uncertainties are expanded using a coverage factor K=2 for a level of confidence of approximately 95%, assuming a normal distribution.

CALIBRATION DATA

Attribute	Nominal	As Found	As Left	Units
Scale Position 1 X	0.2000	0.2001	0.2001	mm
Scale Position 2 X	0.4000	0.4000	0.4000	mm
Scale Position 3 X	0.6000	0.6002	0.6002	mm
Scale Position 4 X	0.8000	0.8001	0.8001	mm
Scale Position 5 X	1.0000	1.0002	1.0002	mm
Scale Position 1 Y	0.2000	0.1999	0.1999	mm
Scale Position 2 Y	0.4000	0.4001	0.4001	mm
Scale Position 3 Y	0.6000	0.6002	0.6002	mm
Scale Position 4 Y	0.8000	0.8002	0.8002	mm
Scale Position 5 Y	1.0000	1.0002	1.0002	mm
Diameter 1.5 mm	1.5000	1.5033	1.5033	mm
Diameter 0.6 mm	0.6000	0.6031	0.6031	mm
Diameter 0.15 mm	0.1500	0.1491	0.1491	mm





Ulrich Métrologie inc.
Ulrich Metrology Inc.
9912, Côte-de-Liesse
Montréal (Québec) H8T 1A1

Tél. (514) 631-6653
Fax (514) 631-6122
info@ulrich.ca
www.ulrich.ca

Certificate no.: 330563	Calibration date: September 19, 2012
Identification: WMT001	Certificate issued: September 19, 2012

CALIBRATION DATA

Attribute	Nominal	As Found	As Left	Units
Diameter 0.07 mm	0.0700	0.0691	0.0691	mm

