

### COMPREHENSIVE TEST REPORT

Customer Name	G & R Cold Forge Inc		SERVICE TICKET/ WORK ORDER #	SV1201130468			
Unit Location	SHIPPING G3		Customer #	N/A			
Mfg/Model #	MT BBA 442		Serial #	2870613-7LH			
Capacity	70	Unit of Measure	16	Number of Divisions	14000	Division Size	0.005

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VP0023IR, and NIST Handbook 44 and/or Canada Standard where applicable.

**SHIFT TEST**

Position	As Found	As Left
1	30.015	30.000
2	30.015	30.000
3	30.015	30.000
4	30.015	30.000

Weights Applied: 3016  
 Max. Permissible Error: ± 4(d)

<input checked="" type="checkbox"/>	In tolerance with <del>out</del> adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 16	0.000	0	0	0.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 16	10.005	+1	± 2	10.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	30 16	30.015	+3	± 4	30.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 16	50.025	+5	± 6	50.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Max. Load*	70 16	70.030	+6	± 8	70.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 16	50.025	+5	± 6	50.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	30 16	30.015	+3	± 4	30.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 16	10.005	+1	± 2	10.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Zero	0 16	0.000	0	± 1	0.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

\* Maximum load used in test  
 No decreasing load applicable

REMARKS:		
WEIGHT IDENTIFICATION NUMBERS:	W15, KIT # 87914	
Weight Traceability Certificate #:	1359699, 1359540	CUSTOMER CALIBRATION DATE: MAR 5, 2012 CUSTOMER CALIBRATION DUE: SEPT/2012
PERFORMED BY:	KARL MILES	<i>[Signature]</i>
	Technician Name (Please print)	Technician Signature
Where Applicable:	Customer Name (Please print)	Customer Signature

### COMPREHENSIVE TEST REPORT

Customer Name	G & R Cold Forging		SERVICE TICKET/ WORK ORDER #	SV1201130468			
Unit Location	SHIPPING B3		Customer #	N/A			
Mfg/Model #	MT 2256		Serial #	1148075-1AJ			
Capacity	5000	Unit of Measure	lb	Number of Divisions	5000	Division Size	1 lb

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VP00231R, and NIST Handbook 44 and/or Canada Standard where applicable.

**SHIFT TEST**

Position	As Found	As Left
1	500	500
2	500	500
3	500	500
4	500	500

Weights Applied: 500 lb  
 Max. Permissible Error: ± 2(d)

<input checked="" type="checkbox"/>	In tolerance without adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 lb	0	0	0	0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
	500 lb	500	0	± 1	500	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
Max. Load*	1000 lb	1000	0	± 2	1000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
	500 lb	500	0	± 1	500	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
Zero	0 lb	0	0	± 1	0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

\* Maximum load used in test  
 No decreasing load applicable

REMARKS:			
WEIGHT IDENTIFICATION NUMBERS:	W11- W30		
Weight Traceability Certificate #:	1359699	CUSTOMER CALIBRATION DATE:	MAR 5, 2012
		CUSTOMER CALIBRATION DUE:	SEP / 2012
PERFORMED BY:	KARL MILES		<i>[Signature]</i>
	Technician Name (Please print)		Technician Signature
Where Applicable:	Customer Name (Please print)		Customer Signature

