

Certificate Report No.: NA6550-278-031213

Mettler Toledo
Service Business Unit Industrial
1900 Polaris Parkway
Columbus, Ohio 43240
1-800-METTLER

METTLER TOLEDO

ISO 9001 Registered

Calibration Certificate

Customer

Company: G & R Cold Forging Inc
Address: 7072 Smith Industrial Drive
City: McGregor *State/Province:* Ontario
Zip/Postal: N0R 1J0 *Customer ID:* C036227001001

Device

Manufacturer: METTLER TOLEDO *Terminal Model:* BBA442
Model: 2256 *Terminal Serial No.:* 2870613-7LH
Serial No.: 1148075-1AJ *Printer Serial No.:* N/A
Capacity: 5000 lb *Dept./Room:* Shipping G3
Readability: 1 lb *No. of Divisions:* 5000
Scale Class: III *Procedure Used:* NIST Handbook 44
Customer Asset #/ID: N/A

Calibration Date: 12-Mar-2013 *Next Cal. Due Date:* 30-Sep-2013
Service Technician: Karl Mikes *Signature:* ELECTRONIC SIGNATURE


Reference Weights

Traceability Test weights used are traceable to the National Institute of Standards and Technology.

| Weight Set No.: | NIST Traceability No.: | Class ASTM/OIML: | Calibration Date: | Calibration Due Date: |
|-----------------|------------------------|------------------|-------------------|-----------------------|
| W1 - W10 | 1359878 | ASTM 5 | 12-Mar-2012 | 12-Mar-2013 |
| 1 - 10 | 1359884 | ASTM 5 | 24-May-2012 | 24-May-2013 |

Measuring Results

Shift Test



| <i>Weights Applied</i> | <i>Position</i> | <i>As Found</i> |
|-----------------------------|-------------------|------------------------|
| | | <i>Displayed Value</i> |
| 1: 500 lb | <i>Position 1</i> | 500 lb |
| 2: 500 lb | <i>Position 2</i> | 500 lb |
| 3: 500 lb | <i>Position 3</i> | 500 lb |
| 4: 500 lb | <i>Position 4</i> | 500 lb |
| <i>Maximum Error:</i> | | 0 lb |
| <i>Max Allowable Error:</i> | | 1 lb |

Linearity

| | <i>As Found</i> | | | | | |
|--------|------------------------|----------------|--------------|-----|------------------------|-------------------|
| | <i>Weights Applied</i> | <i>Reading</i> | <i>Error</i> | | <i>Allowable Error</i> | <i>Within Tol</i> |
| Zero 1 | 0 lb | 0 lb | 0 lb | 0 d | 1 d | YES |
| 2 | 500 lb | 500 lb | 0 lb | 0 d | 1 d | YES |
| Max 3 | 1000 lb | 1000 lb | 0 lb | 0 d | 2 d | YES |
| 4 | 500 lb | 500 lb | 0 lb | 0 d | 1 d | YES |
| Zero 5 | 0 lb | 0 lb | 0 lb | 0 d | 1 d | YES |

Substitution method used

Adjustment of scale has been necessary

In case of "NO", "AS FOUND" results correspond to "AS LEFT"

YES

NO

Remarks

None.