

Certificate Report No.: NA6550-280-031213

Mettler Toledo
Service Business Unit Industrial
1900 Polaris Parkway
Columbus, Ohio 43240
1-800-METTLER

METTLER TOLEDO

ISO 9001 Registered

Calibration Certificate

Customer

Company: G & R Cold Forging Inc
Address: 7072 Smith Industrial Drive
City: McGregor *State/Province:* Ontario
Zip/Postal: N0R 1J0 *Customer ID:* C036227001001

Device

Manufacturer: METTLER TOLEDO *Terminal Model:* IND445-RD
Model: CB60L *Terminal Serial No.:* 2791511
Serial No.: 16490786KF *Printer Serial No.:* N/A
Capacity: 100 lb *Dept./Room:* Shipping G4
Readability: 0.01 lb *No. of Divisions:* 10000
Scale Class: III *Procedure Used:* NIST Handbook 44
Customer Asset #/ID: N/A

Calibration Date: 12-Mar-2013 *Next Cal. Due Date:* 30-Sep-2013
Service Technician: Karl Mikes *Signature:* ELECTRONIC SIGNATURE


Reference Weights

Traceability Test weights used are traceable to the National Institute of Standards and Technology.

Weight Set No.:	NIST Traceability No.:	Class ASTM/OIML:	Calibration Date:	Calibration Due Date:
W1 - W10	1359878	ASTM 5	12-Mar-2012	12-Mar-2013
Kit # 87914	1359858	ASTM 5	28-Aug-2012	28-Aug-2013

Measuring Results

Shift Test



<i>Weights Applied</i>	<i>Position</i>	<i>As Found</i>
		<i>Displayed Value</i>
1: 25.00 lb	<i>Position 1</i>	25.00 lb
2: 25.00 lb	<i>Position 2</i>	25.00 lb
3: 25.00 lb	<i>Position 3</i>	25.00 lb
4: 25.00 lb	<i>Position 4</i>	25.00 lb
<i>Maximum Error:</i>		0.00 lb
<i>Max Allowable Error:</i>		0.03 lb

Linearity

	<i>As Found</i>					
	<i>Weights Applied</i>	<i>Reading</i>	<i>Error</i>		<i>Allowable Error</i>	<i>Within Tol</i>
Zero 1	0.00 lb	0.00 lb	0.00 lb	0 d	1 d	YES
2	50.00 lb	50.00 lb	0.00 lb	0 d	5 d	YES
Max 3	100.00 lb	100.00 lb	0.00 lb	0 d	5 d	YES
4	50.00 lb	50.00 lb	0.00 lb	0 d	5 d	YES
Zero 5	0.00 lb	0.00 lb	0.00 lb	0 d	1 d	YES

Substitution method used

Adjustment of scale has been necessary

In case of "NO", "AS FOUND" results correspond to "AS LEFT"

YES

NO

Remarks

None.