




**ABSOLUTE MEASUREMENT SYSTEMS**  
**CMM Calibration & Service**  
 3986 Brantidge Road  
 Mississauga, ON L5N 7V7  
 Phone: (905) 785-7600



**LABORATORY ACCREDITATION BUREAU**  
 ACCREDITED ISO/IEC 17025  
 Certificate # L2196 Calibration

# Calibration Certificate

Certificate Number: **C-0172**

Reissue Number: \_\_\_\_\_  
 if applicable

Company: **Windsor Machine & Stamping (2009) Ltd.**  
 Address: **7084 Smith Industrial Drive**  
 City: **McGragor**  
 Province: **Ontario**  
 Postal Code: **N0P 1J0**

CMM: **IMS Impact II**  
 CMM Software: **Virtual DMS**  
 Serial No.: **600072**  
 Gauge No.: \_\_\_\_\_  
 Technician: **Robert Scharff**

Procedure: **B89.4.1b-2001**  
 Units: **Millimeters**  
 Temp. Range: **22.7 °C - Max 21.3 °C - Min**  
 Date Calibrated: **March 1 2012**  
 Date Issued: **April 19 2012**

Repeatability		
Specification	As Found	After Correction
X	N/A	0.001
Y	N/A	0.001
Z	N/A	0.002
		Uncertainty 1.2 µm 1.2 µm 1.2 µm

Linear Accuracy		
Specification	As Found	After Correction
X	N/A	0.004
Y	N/A	0.016
Z	N/A	0.001
		Uncertainty (0.3 + 2.5L) µm (0.3 + 2.5L) µm (0.3 + 2.5L) µm

Linear accuracy is measured in 1 m increments as per AMS procedure.

Squareness**		
Specification	As Found	After Correction
XY	N/A	0.0001 degrees
YZ	N/A	-0.0106 degrees
ZX	N/A	0.0036 degrees
		After Correction 0.0001 degrees -0.0001 degrees 0.0001 degrees
		Uncertainty (0.8 + 1.5L) µm (0.8 + 1.5L) µm (0.8 + 1.5L) µm

Volumetric Performance		
Specification	As Found	After Correction
N/A	0.073	0.040
		Uncertainty (0.8 + 1.5L) µm

This Calibration Certificate certifies that the instrument listed above has been calibrated in accordance with applicable national, international, and AMS specifications and standards. This Certificate of Calibration is issued in accordance with the applicable requirements of ISO/IEC 17025:2005. Calibration results are traceable to the National Research Council (NRC), the National Institute of Standards and Technology (NIST), or the National Physical Laboratory (NPL), where applicable. The above noted CMM was calibrated at the Customer's location. The calibration results relate only to the CMM noted above, and for the environmental and instrument conditions at the time of calibration. This Calibration Certificate is part of a multi-page document containing 8 pages.

**Traceability Information:**

Length Reference Standard: **Optodyne Laser (MCV 500) - S.N. 0807003080, S.N. 0807002807**  
 Environmental Thermometer: **Fluke Digital Thermometer - S.N. 97890209**  
 Ballbar Spheres: **Renishaw 25.000 mm - S.N. 138481, S.N. 138485**  
 Ball bar length: **mm**

\*After Correction results are only reported for coordinate measuring machines whose volumetric parameters have been adjusted.  
 \*\* Squareness measurements are not part of AMS' scope of accreditation, but are taken for reference purposes (if requested).

\*The results relate only to this CMM, environmental and instrument conditions at the time of calibration.  
 \*\*This certificate may only be reproduced in its entirety.

**Measurement Uncertainty:**

The stated expanded uncertainties are given at a coverage factor of k=2 for a level of confidence of approximately 95% assuming a normal distribution.

Name: **Robert Scharff**  
 Title: **Service Manager**

Signature: \_\_\_\_\_

**Repeatability 'As Found'**

Meas. #	X Axis	Y Axis	Z Axis
1	-0.0006	0.0018	-0.0032
2	-0.0010	0.0020	-0.0032
3	-0.0007	0.0015	-0.0028
4	-0.0010	0.0021	-0.0034
5	-0.0012	0.0019	-0.0035
6	-0.0012	0.0017	-0.0030
7	-0.0018	0.0018	-0.0031
8	-0.0017	0.0017	-0.0033
9	-0.0011	0.0012	-0.0026
10	-0.0010	0.0013	-0.0031

Max: -0.0006  
Min: -0.0018  
Repeatability: 0.0012

**Volumetric Performance 'As Found'**

Position	Trial 1	Trial 2	Trial 3	Average
1	394.048	N/A	N/A	394.048
2	394.067	N/A	N/A	394.067
3	394.051	N/A	N/A	394.052
4	394.052	N/A	N/A	394.052
5	394.050	N/A	N/A	394.050
6	394.078	N/A	N/A	394.078
7	394.053	N/A	N/A	394.053
8	394.064	N/A	N/A	394.064
9	394.074	N/A	N/A	394.074
10	N/A	N/A	N/A	N/A
11	N/A	N/A	N/A	N/A
12	N/A	N/A	N/A	N/A
13	N/A	N/A	N/A	N/A
14	N/A	N/A	N/A	N/A
15	N/A	N/A	N/A	N/A
16	N/A	N/A	N/A	N/A
17	N/A	N/A	N/A	N/A
18	N/A	N/A	N/A	N/A
19	N/A	N/A	N/A	N/A
20	N/A	N/A	N/A	N/A
21	N/A	N/A	N/A	N/A
22	N/A	N/A	N/A	N/A
23	N/A	N/A	N/A	N/A
24	N/A	N/A	N/A	N/A
25	N/A	N/A	N/A	N/A
26	N/A	N/A	N/A	N/A
27	N/A	N/A	N/A	N/A
28	N/A	N/A	N/A	N/A
29	N/A	N/A	N/A	N/A
30	N/A	N/A	N/A	N/A
31	N/A	N/A	N/A	N/A
32	N/A	N/A	N/A	N/A
33	N/A	N/A	N/A	N/A
34	N/A	N/A	N/A	N/A
35	N/A	N/A	N/A	N/A

Max: 394.099  
Min: 394.026  
Volumetric: 0.073

Units: Millimeters

**Repeatability 'After Correction'**

Meas. #	X Axis	Y Axis	Z Axis
1	-0.0026	0.0010	-0.0039
2	-0.0031	0.0005	-0.0022
3	-0.0036	0.0013	-0.0022
4	-0.0028	0.0003	-0.0019
5	-0.0040	0.0007	-0.0022
6	-0.0038	0.0002	-0.0020
7	-0.0039	0.0008	-0.0020
8	-0.0034	0.0008	-0.0028
9	-0.0037	0.0005	-0.0033
10	-0.0037	0.0008	-0.0024

Max: -0.0026  
Min: -0.0040  
Repeatability: 0.0014

**Volumetric Performance 'After Correction'**

Position	Trial 1	Trial 2	Trial 3	Average
1	394.084	N/A	N/A	394.084
2	394.066	N/A	N/A	394.066
3	394.052	N/A	N/A	394.052
4	394.044	N/A	N/A	394.044
5	394.058	N/A	N/A	394.058
6	394.074	N/A	N/A	394.074
7	394.055	N/A	N/A	394.055
8	394.055	N/A	N/A	394.055
9	394.082	N/A	N/A	394.082
10	N/A	N/A	N/A	N/A
11	N/A	N/A	N/A	N/A
12	N/A	N/A	N/A	N/A
13	N/A	N/A	N/A	N/A
14	N/A	N/A	N/A	N/A
15	N/A	N/A	N/A	N/A
16	N/A	N/A	N/A	N/A
17	N/A	N/A	N/A	N/A
18	N/A	N/A	N/A	N/A
19	N/A	N/A	N/A	N/A
20	N/A	N/A	N/A	N/A
21	N/A	N/A	N/A	N/A
22	N/A	N/A	N/A	N/A
23	N/A	N/A	N/A	N/A
24	N/A	N/A	N/A	N/A
25	N/A	N/A	N/A	N/A
26	N/A	N/A	N/A	N/A
27	N/A	N/A	N/A	N/A
28	N/A	N/A	N/A	N/A
29	N/A	N/A	N/A	N/A
30	N/A	N/A	N/A	N/A
31	N/A	N/A	N/A	N/A
32	N/A	N/A	N/A	N/A
33	N/A	N/A	N/A	N/A
34	N/A	N/A	N/A	N/A
35	N/A	N/A	N/A	N/A

Max: 394.084  
Min: 394.044  
Volumetric: 0.040

**Squareness\*\* 'As Found'**

	XY	YZ	ZX
Trial 1	+X -Y 394.061	+Y -Z 394.056	+Z -X 394.074
Trial 2	+X -Y N/A	+Y -Z N/A	+Z -X N/A
Trial 3	+X -Y N/A	+Y -Z N/A	+Z -X N/A
Average:	394.061	394.056	394.074
Angle:	-0.0007 degrees	-0.0106 degrees	0.0036 degrees
Squareness:	1.57081 arc sec.	1.57098 arc sec.	1.57073 arc sec.
Range:	0.0005	0.0773	-0.025

**Squareness\*\* 'After Correction'**

	XY	YZ	ZX
Trial 1	+X -Y 394.063	+Y -Z 394.067	+Z -X 394.062
Trial 2	+X -Y 394.062	+Y -Z 394.064	+Z -X 394.066
Trial 3	+X -Y 394.063	+Y -Z 394.067	+Z -X 394.062
Average:	394.063	394.064	394.066
Angle:	0.0001 degrees	-0.0001 degrees	0.0001 degrees
Squareness:	1.57079 arc sec.	1.57080 arc sec.	1.57079 arc sec.
Range:	-0.001	0.001	-0.001



Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Reissue Number: \_\_\_\_\_  
if applicable

Date Calibrated: March 1 2012

**Linear Displacement - As Found**

X Axis

Cell	Location	Trial 1	Trial 2	Trial 3	Average
1	0	0.000	0.000	0.000	0.000
2	75	0.002	0.004	0.004	0.003
3	150	0.002	0.003	0.003	0.003
4	225	0.002	0.004	0.005	0.004
5	300	0.001	0.002	0.004	0.002
6	375	0.002	0.005	0.003	0.003
7	450	0.001	0.003	0.005	0.003
8	525	0.001	0.003	0.004	0.003
9	600	0.000	0.002	0.003	0.002
10	675	0.001	0.003	0.003	0.002
11	750	0.002	0.003	0.007	0.004

Range								
Axis	Trial 1		Trial 2		Trial 3		Average	
	Max	Min	Max	Min	Max	Min	Max	Min
X Axis	0.002	0.000	0.005	0.000	0.007	0.000	0.004	0.000

Linear Displacement Summary	
As Found	
Axis	Result
X Axis	0.004



Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Reissue Number: \_\_\_\_\_  
if applicable

Date Calibrated: March 1 2012

**Linear Displacement - As Found**

Y Axis	Cell	Location	Trial 1	Trial 2	Trial 3	Average
	1	0	0.000	0.000	0.000	0.000
	2	75	0.001	0.001	0.002	0.001
	3	150	0.003	0.002	0.003	0.003
	4	225	0.005	0.004	0.005	0.005
	5	300	0.007	0.006	0.007	0.007
	6	375	0.007	0.007	0.008	0.007
	7	450	0.009	0.009	0.009	0.009
	8	525	0.010	0.010	0.010	0.010
	9	600	0.012	0.013	0.013	0.013
	10	675	0.015	0.015	0.015	0.015
	11	750	0.016	0.016	0.017	0.016

Range								
Axis	Trial 1		Trial 2		Trial 3		Average	
	Max	Min	Max	Min	Max	Min	Max	Min
Y Axis	0.016	0.000	0.016	0.000	0.017	0.000	0.016	0.000

Linear Displacement Summary	
As Found	
Axis	Result
Y Axis	0.016



**ABSOLUTE MEASUREMENT SYSTEMS**  
CMM Calibration & Service

Certificate Number: C-0172

Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Reissue Number: \_\_\_\_\_  
if applicable

Date Calibrated: March 1 2012

**Linear Displacement - As Found**

Z Axis

Cell	Location	Trial 1	Trial 2	Trial 3	Average
1	0	0.000	0.000	0.000	0.000
2	55	0.000	0.000	0.000	0.000
3	110	0.000	0.000	0.000	0.000
4	165	0.000	0.000	0.000	0.000
5	220	0.000	0.000	0.000	0.000
6	275	0.001	0.000	0.000	0.000
7	330	0.000	-0.001	-0.001	-0.001
8	385	0.001	0.000	0.000	0.000
9	440	0.001	0.000	-0.001	0.000
10	495	0.001	0.000	0.000	0.000

Range								
Axis	Trial 1		Trial 2		Trial 3		Average	
	Max	Min	Max	Min	Max	Min	Max	Min
Z Axis	0.001	0.000	0.000	-0.001	0.000	-0.001	0.000	-0.001

Linear Displacement Summary	
As Found	
Axis	Result
Z Axis	0.001

\*The results relate only to this CMM, environmental and instrument conditions at the time of calibration.  
\*This certificate may only be reproduced in its entirety.



Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Reissue Number: \_\_\_\_\_  
if applicable

Date Calibrated: March 1 2012

**Linear Displacement - After Correction**

X Axis	Cell	Location	Trial 1	Trial 2	Trial 3	Average
	1	0	0.000	0.000	0.000	0.000
	2	75	0.001	0.001	0.001	0.001
	3	150	0.001	0.000	0.001	0.001
	4	225	0.001	0.001	0.001	0.001
	5	300	0.000	0.001	0.000	0.000
	6	375	0.001	0.001	0.001	0.001
	7	450	0.001	0.001	0.001	0.001
	8	525	0.001	0.001	0.001	0.001
	9	600	0.000	0.000	0.000	0.000
	10	675	0.000	0.000	0.000	0.000
	11	750	0.001	0.001	0.001	0.001

Range								
Axis	Trial 1		Trial 2		Trial 3		Average	
	Max	Min	Max	Min	Max	Min	Max	Min
X Axis	0.001	0.000	0.001	0.000	0.001	0.000	0.001	0.000

Linear Displacement Summary	
After Correction	
Axis	Result
X Axis	0.001



Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Reissue Number: \_\_\_\_\_  
if applicable

Date Calibrated: March 1 2012

**Linear Displacement - After Correction**

Y Axis

Cell	Location	Trial 1	Trial 2	Trial 3	Average
1	0	0.000	0.000	0.000	0.000
2	75	0.000	0.001	0.000	0.000
3	150	0.000	0.000	0.000	0.000
4	225	0.000	0.000	0.000	0.000
5	300	0.000	0.001	0.000	0.000
6	375	0.001	0.001	0.001	0.001
7	450	0.000	0.001	0.000	0.000
8	525	0.000	0.000	0.000	0.000
9	600	0.000	0.000	0.000	0.000
10	675	0.000	0.000	0.000	0.000
11	750	0.001	0.001	0.001	0.001

Range								
Axis	Trial 1		Trial 2		Trial 3		Average	
Y Axis	Max	Min	Max	Min	Max	Min	Max	Min
	0.001	0.000	0.001	0.000	0.001	0.000	0.001	0.000

Linear Displacement Summary	
After Correction	
Axis	Result
Y Axis	0.001



Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Reissue Number: \_\_\_\_\_  
 if applicable

Date Calibrated: March 1 2012

**Linear Displacement - After Correction**

Z Axis	Cell	Location	Trial 1	Trial 2	Trial 3	Average
	1	0				
	2	55				
	3	110				
	4	165				
	5	220				
	6	275				
	7	330				
	8	385				
	9	440				
	10	495				

Range								
Axis	Trial 1		Trial 2		Trial 3		Average	
Z Axis	Max	Min	Max	Min	Max	Min	Max	Min
	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A

Linear Displacement Summary	
After Correction	
Axis	Result
Z Axis	N/A