



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

FRANK COX METROLOGY
(Formerly CANADIAN CENTRAL GAUGE LABORATORY)
40 West Drive
Brampton, Ontario, Canada L6T 3T6
Hilliard Cox Phone: 905 457 9190

CALIBRATION

Valid To: January 31, 2015

Certificate Number: 1165.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1,4}:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2,5} (\pm)	Comments
Gage Blocks – Steel, Ceramic, Carbide, Chrome Carbide Length Flatness Parallelism Length Flatness Parallelism	(0.5 to 100) mm (0.010 to 4) in	(0.075 + 0.025L) μ m 0.1 μ m 0.05 μ m (3 + 1L) μ in 4 μ in 2 μ in	ANSI B89.1.9
Plain Plug Gages – Thread Measure Wire	(0.007 to 4) in (>4 to 24) in Up to 80 TPI	(4 + 1.5L) μ in (10 + 3L) μ in 4 μ in	ANSI B89.1.5 ANSI B891.17
Plain Ring Gages	(0.04 to 6) in (6 to 12) in	(7 + 1.5L) μ in (15 + 1.5L) μ in	ANSI B89.1.6M

Peter Abney
Page 1 of 7

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
Micrometers – Outside ³ Height Setting ³ High Resolution	Up to 1 in (2 to 36) in (6 to 24) in Up to 4 in	$(0.6R + 10L) \mu\text{in}$ $(0.6R + 5L) \mu\text{in}$ $(0.6R + 10L) \mu\text{in}$ 10 μin	By comparison
Depth Gages ³	(1 to 24) in	$(0.6R + 10L) \mu\text{in}$	By comparison
Squares	18 in	22 $\mu\text{in/in}$	Square checker
Thread Plug Gages – Simple Pitch Diameter Linear Pitch Variation Flank Angle	Up to 4 in (4 to 12) in Up to 16 in Up to 16 in	85 μin $(85 + 5L) \mu\text{in}$ 35 μin per 4 in 3 min of arc depending on pitch	By comparison Helical path analyzer or ULM Optical comparator
Thread Ring Gages – Adjustable Type Functional Pitch Diameter Flank Angle	(0.06 to 12) in (0.06 to 12) in	$(320 + 15L) \mu\text{in}$ 3 min of arc depending on pitch	By comparison, fit to master plug By cast method

Parameter/Equipment	Range	CMC ^{2,5} (\pm)	Comments
Thread Ring Gages – Non-Adjustable Type			
Simple Pitch Diameter	(0.5 to 6) in	120 μ in	By comparison
Lead Variation	(0.5 to 2) in	25 μ in per 4 in	
Flank Angle	(0.5 to 2) in	3 min of arc depending on pitch	By cast method
Thread Caliper Gauges – Adjustable			
Knife Edge	Up to 12 in	(420 + 5L) μ in	Fit to master
Roller Type	Up to 12 in	(250 + 10L) μ in	
Vernier, Dial, and Digital Calipers ³	Up to 60 in	(0.6R + 4.5L) μ in	By comparison
Dial, Digital, and Test Indicators –	Up to 2 in	30 μ in	ANSI B89.1.10M
High-Resolution/Digital	(0.001 to 0.05) in	1.2 μ in	
High-Resolution/Analog	Up to 0.1 in	4.5 μ in	
Bore Gages	(0.5 to 24) in	(0.6R + 3L) μ in	By comparison
Length Standards – Setting Rods	(1 to 36) in	(6 + 4.5L) μ in	By comparison
Surface Plates ³ - Repeat Reading Flatness	0.020 in Up to 20 ft diagonal	4.5 μ in (50 + 4D) μ in	GGG-P-463c; D is the length of the diagonal in feet.

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
Optical Comparators ³ – Magnification Linearity Angle	Up to 100x 18 in travel Up to 360°	Lens: < 1% magnification 200 μin Angle: 2.7 min	Opti-master, mag rule, angle blocks, balls
High Resolution Comparators	2 in	1.5 μin	Master blocks
Digital, Dial, and Vernier Height Gauges ³	(6 to 48) in	(0.6R + 5.2L) μin	By comparison
Sine Bars and Plates	Up to 10 in	30 μin/5 in	By comparison
Precision Levels	Up to 20 in	(0.6R + 5.5) μin	Level test rig
Autocollimator	5 min	0.31 arc sec	Gauge blocks and sine equipment
Plain Snap Gages ³	Up to 12 in	(120 + 3L) μin	By comparison
Steel Rules	Up to 24 in	0.003 in	Optical comparator

II. Mechanical

Parameter	Range	CMC ² (±)	Comments
Indirect Verification of Rockwell Hardness Testers ³	HRA (25 to 35) HRA (36 to 59) HRA (60 to 85) HRA	0.43 HRA 0.54 HRA 0.35 HRA	Indirect verification method per ASTM E18

Parameter	Range	CMC ^{2,5} (±)	Comments
Indirect Verification of Rockwell Hardness Testers ³ – (cont)	HRBW (40 to 59) HRB (60 to 79) HRB (80 to 100) HRB HRC (20 to 30) HRC (31 to 55) HRC (56 to 65) HRC HREW (55 to 75) HREW (76 to 90) HREW (91 to 115) HREW HR15N (35 to 65) HR15N (66 to 75) HR15N (76 to 95) HR15N HR15TW (45 to 69) HR15TW (70 to 84) HR15TW (85 to 95) HR15TW HR30N (40 to 50) HR30N (55 to 74) HR30N (75 to 85) HR30N HR30TW (25 to 49) HR30TW (50 to 70) HR30TW (71 to 85) HR30TW HR45N (35 to 49) HR45N (50 to 65) HR45N (66 to 75) HR45N HR45TW (30 to 55) HR45TW (56 to 70) HR45TW (71 to 75) HR45TW	1.2 HRB 0.81 HRB 0.55 HRB 0.49 HRC 0.74 HRC 0.38 HRC 0.67 HREW 0.75 HREW 0.56 HREW 0.49 HR15N 0.71 HR15N 0.32 HR15N 0.61 HR15TW 0.47 HR15TW 0.58 HR15TW 0.49 HR30N 0.99 HR30N 0.45 HR30N 0.74 HR30TW 0.45 HR30TW 0.50 HR30TW 0.66 HR45N 0.82 HR45N 0.44 HR45N 0.77 HR45TW 0.52 HR45TW 0.57 HR45TW	Indirect verification method per ASTM E18
Torque Tools ³	450 ft·lbf	1.0 % IV	Torque tester
Torque Testers	5500 in·lbf	0.16 % IV	Standard weights

Parameter	Range	CMC ^{2,5} (±)	Comments
Force Gauges	250 lbf	1.5 % IV	By comparison with standard weights
Direct Verification of Durometers –	Shore Types A, B, C, D, DO, M, O, and OO		ASTM D 2240
Indenter Shape and Extension:			
Extension at Zero	---	0.0003 in	
35° Conical Frustum Indentor	Diameter at frustum base	0.0003 in	
	Diameter at top of frustum	0.0003 in	
	Cone angle	8 arcmin	
30° Cone Indentor	Diameter at base of cone	0.0003 in	
Indenter Shape and Extension:			
1.2 mm Radius Indentor	Cone angle	8 arcmin	
	Tip Radius	0.0003 in	
	Indentor diameter	0.0002 in	
	Indentor radius	0.0003 in	
Verification of the Durometer Spring	---	1.6 grams	Durocalibrator

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ This laboratory offers metric equivalent capability for all items listed.

⁵ In the statement of CMC, L is the numerical value of the nominal length of the device measured in inches; R is the numerical value of the resolution of the device in microinches; D is the numerical value of the nominal diameter of the device measured in inches except where noted; IV is the percent of indicated value.



American Association for Laboratory Accreditation

Accredited Laboratory

A2LA has accredited

FRANK COX METROLOGY

Brampton, Ontario, Canada

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF *Communiqué* dated 8 January 2009).

Presented this 25th day of February 2013.





President & CEO

For the Accreditation Council
Certificate Number 1165.01
Valid to January 31, 2015

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.